


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX013020\
Data File : VX014767.D
Acq On : 30 Jan 2020 09:46
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
Title : SW846 8260
Last Update : Wed Jan 22 12:35:15 2020



AutoFind: Scans 1679, 1680, 1681; Background Corrected with Scan 1672

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.0	16327	PASS
75	95	30	60	49.7	40545	PASS
95	95	100	100	100.0	81618	PASS
96	95	5	9	6.5	5298	PASS
173	174	0.00	2	1.7	1117	PASS
174	95	50	100	82.0	66955	PASS
175	174	5	9	8.1	5414	PASS
176	174	95	101	98.2	65739	PASS
177	176	5	9	7.1	4670	PASS