

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013021\
 Data File : VX020808.D
 Acq On : 29 Jan 2021 17:34
 Operator : JC/SP
 Sample : M1285-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 30 02:37:15 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	293338	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	495187	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	469591	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	198789	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	197380	49.33	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.66%
35) Dibromofluoromethane	5.464	113	167894	49.50	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.00%
50) Toluene-d8	8.696	98	606810	48.80	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.60%
62) 4-Bromofluorobenzene	11.122	95	212659	44.04	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.08%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013021\
Data File : VX020808.D
Acq On : 29 Jan 2021 17:34
Operator : JC/SP
Sample : M1285-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 30 02:37:15 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

