

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013021\
 Data File : VX020814.D
 Acq On : 29 Jan 2021 19:55
 Operator : JC/SP
 Sample : M1296-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW6-GW-98-100

Quant Time: Jan 30 02:39:02 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	288783	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	498025	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.097	117	485373	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.060	152	211007	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	198619	50.42	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.84%
35) Dibromofluoromethane	5.464	113	171611	50.31	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.62%
50) Toluene-d8	8.695	98	606302	48.48	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.96%
62) 4-Bromofluorobenzene	11.122	95	216155	44.50	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.00%
Target Compounds						
16) Acetone	2.416	43	4960	3.32	ug/l	98
64) Tetrachloroethene	9.311	164	916	0.25	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013021\
Data File : VX020814.D
Acq On : 29 Jan 2021 19:55
Operator : JC/SP
Sample : M1296-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW6-GW-98-100

Quant Time: Jan 30 02:39:02 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

