

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013021\
 Data File : VX020815.D
 Acq On : 29 Jan 2021 20:18
 Operator : JC/SP
 Sample : M1296-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW6-GW-158-160

Quant Time: Feb 01 04:52:22 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	283972	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	484186	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	463103	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	207192	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	195238	50.41	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.82%	
35) Dibromofluoromethane	5.464	113	167895	50.62	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.24%	
50) Toluene-d8	8.695	98	599454	49.30	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.60%	
62) 4-Bromofluorobenzene	11.122	95	209631	44.39	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 88.78%	
Target Compounds						
						Qvalue
16) Acetone	2.416	43	4394	2.99	ug/l	95
30) Chloroform	5.184	83	5520	0.89	ug/l	95
44) Trichloroethene	7.190	130	34574	9.08	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013021\
Data File : VX020815.D
Acq On : 29 Jan 2021 20:18
Operator : JC/SP
Sample : M1296-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW6-GW-158-160

Quant Time: Feb 01 04:52:22 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

