

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013021\
 Data File : VX020817.D
 Acq On : 29 Jan 2021 21:05
 Operator : JC/SP
 Sample : M1296-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW6-GW-218-220

Quant Time: Jan 30 02:39:55 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	281237	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	478365	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	462503	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	205622	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	193515	50.45	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.90%	
35) Dibromofluoromethane	5.464	113	166605	50.85	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.70%	
50) Toluene-d8	8.695	98	596518	49.66	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 99.32%	
62) 4-Bromofluorobenzene	11.122	95	211023	45.23	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 90.46%	
Target Compounds						
						Qvalue
16) Acetone	2.416	43	3415	2.35	ug/l	94
30) Chloroform	5.172	83	2881	0.47	ug/l	92
44) Trichloroethene	7.184	130	26735	7.11	ug/l	93
64) Tetrachloroethene	9.311	164	4612	1.30	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013021\
Data File : VX020817.D
Acq On : 29 Jan 2021 21:05
Operator : JC/SP
Sample : M1296-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW6-GW-218-220

Quant Time: Jan 30 02:39:55 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

