

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013021\
 Data File : VX020818.D
 Acq On : 29 Jan 2021 21:28
 Operator : JC/SP
 Sample : M1296-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW6-GW-238-240

Manual Integrations
 APPROVED

MMDadoda
 2/3/2021 9:54:03 AM

Quant Time: Jan 30 02:40:14 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	279225	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	469824	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	436447	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	174820	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	189479	49.75	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.50%
35) Dibromofluoromethane	5.458	113	164972	51.26	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.52%
50) Toluene-d8	8.696	98	579434	49.11	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.22%
62) 4-Bromofluorobenzene	11.122	95	193391	42.21	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	84.42%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.185	85	1703	1.58	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.367	101	11863	4.37	ug/l	94
12) 1,1-Dichloroethene	2.355	96	2465	0.92	ug/l	98
16) Acetone	2.416	43	3350	2.32	ug/l #	80
24) 1,1-Dichloroethane	3.672	63	3544	0.60	ug/l #	92
27) cis-1,2-Dichloroethene	4.568	96	3559	0.97	ug/l	97
30) Chloroform	5.178	83	3389	0.56	ug/l	94
32) 1,1,1-Trichloroethane	5.452	97	1625m	0.30	ug/l	
44) Trichloroethene	7.184	130	277107	74.99	ug/l	99
64) Tetrachloroethene	9.317	164	5579	1.67	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013021\
Data File : VX020818.D
Acq On : 29 Jan 2021 21:28
Operator : JC/SP
Sample : M1296-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW6-GW-238-240

Manual Integrations
APPROVED

MMDadoda
2/3/2021 9:54:03 AM

Quant Time: Jan 30 02:40:14 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

