

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013021\
 Data File : VX020819.D
 Acq On : 29 Jan 2021 21:52
 Operator : JC/SP
 Sample : M1291-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 2021-01-27-EB

Quant Time: Jan 30 02:40:34 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	276055	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	475023	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	449752	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	187036	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.032	65	190972	50.72	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.44%
35) Dibromofluoromethane	5.465	113	164860	50.67	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.34%
50) Toluene-d8	8.696	98	592411	49.66	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.32%
62) 4-Bromofluorobenzene	11.122	95	202146	43.64	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.28%
Target Compounds						
					Qvalue	
16) Acetone	2.416	43	56927	39.88	ug/l	97
18) Methyl Acetate	2.752	43	2214	0.61	ug/l #	83
25) 2-Butanone	4.642	43	6874	3.07	ug/l	95
37) Ethyl Acetate	4.806	43	2733	0.60	ug/l #	92
44) Trichloroethene	7.184	130	925	0.25	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013021\
Data File : VX020819.D
Acq On : 29 Jan 2021 21:52
Operator : JC/SP
Sample : M1291-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
2021-01-27-EB

Quant Time: Jan 30 02:40:34 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

