



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013021\
Data File : VX020801.D
Acq On : 29 Jan 2021 13:04
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Title : SW846 8260
Last Update : Mon Jan 25 13:38:15 2021



AutoFind: Scans 1644, 1645, 1646; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.1	5276	PASS
75	95	30	60	51.8	15973	PASS
95	95	100	100	100.0	30859	PASS
96	95	5	9	7.2	2212	PASS
173	174	0.00	2	1.1	253	PASS
174	95	50	100	76.9	23725	PASS
175	174	5	9	7.0	1652	PASS
176	174	95	101	95.0	22540	PASS
177	176	5	9	6.5	1455	PASS