

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013023\
 Data File : VX033946.D
 Acq On : 30 Jan 2023 15:05
 Operator : JC/MD
 Sample : 01361-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10-EB-20230125

Quant Time: Jan 31 04:46:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	162642	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	234490	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208400	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98621	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77616	45.324	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.640%
35) Dibromofluoromethane	5.385	113	71851	49.346	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.700%
50) Toluene-d8	8.646	98	265796	50.101	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.200%
62) 4-Bromofluorobenzene	11.079	95	90839	46.118	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013023\
Data File : VX033946.D
Acq On : 30 Jan 2023 15:05
Operator : JC/MD
Sample : 01361-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10-EB-20230125

Quant Time: Jan 31 04:46:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

