

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013023\
 Data File : VX033947.D
 Acq On : 30 Jan 2023 15:28
 Operator : JC/MD
 Sample : 01356-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-J

Quant Time: Jan 31 04:47:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	162682	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	234118	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	204022	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93572	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	77806	45.424	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.840%
35) Dibromofluoromethane	5.385	113	71633	49.275	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.540%
50) Toluene-d8	8.647	98	267801	50.559	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.120%
62) 4-Bromofluorobenzene	11.079	95	88456	44.979	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.960%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	15846	13.748	ug/l	95
20) Methylene Chloride	2.788	84	7351	2.559	ug/l	88
37) Ethyl Acetate	4.715	43	19003	6.925	ug/l	98
43) Isopropyl Acetate	6.336	43	81419	24.308	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013023\
Data File : VX033947.D
Acq On : 30 Jan 2023 15:28
Operator : JC/MD
Sample : 01356-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-J

Quant Time: Jan 31 04:47:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

