

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013023\
 Data File : VX033954.D
 Acq On : 30 Jan 2023 18:23
 Operator : JC/MD
 Sample : 01328-25
 Misc : 13.53g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOTAL-VOC-BORING-2-0-7-GRAB-7

Quant Time: Jan 31 04:51:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	152988	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	215904	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186139	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95135	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	70697	43.889	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.780%
35) Dibromofluoromethane	5.385	113	61437	45.826	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.660%
50) Toluene-d8	8.647	98	241852	49.512	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.020%
62) 4-Bromofluorobenzene	11.085	95	85120	46.934	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013023\
Data File : VX033954.D
Acq On : 30 Jan 2023 18:23
Operator : JC/MD
Sample : 01328-25
Misc : 13.53g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOTAL-VOC-BORING-2-0-7-GRAB-7

Quant Time: Jan 31 04:51:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

