

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013023\
 Data File : VX033955.D
 Acq On : 30 Jan 2023 18:47
 Operator : JC/MD
 Sample : 01328-05
 Misc : 5.21g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOTAL-VOC-1-BORING-1-0-7-GRAB-2

Quant Time: Jan 31 04:52:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	159446	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	221402	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	190272	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93237	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72138	42.969	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.940%
35) Dibromofluoromethane	5.385	113	63830	46.429	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.860%
50) Toluene-d8	8.647	98	248356	49.581	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.160%
62) 4-Bromofluorobenzene	11.085	95	83456	44.874	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013023\
Data File : VX033955.D
Acq On : 30 Jan 2023 18:47
Operator : JC/MD
Sample : 01328-05
Misc : 5.21g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOTAL-VOC-1-BORING-1-0-7-GRAB-2

Quant Time: Jan 31 04:52:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

