

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013023\
 Data File : VX033956.D
 Acq On : 30 Jan 2023 19:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 31 04:52:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	0.395	0.334	15.4	73	0.00
3 P	Chloromethane	0.443	0.406	8.4	72	0.00
4 C	Vinyl Chloride	0.424	0.419	1.2#	76	0.00
5 T	Bromomethane	0.243	0.198	18.5	75	0.00
6 T	Chloroethane	0.262	0.253	3.4	83	0.00
7 T	Trichlorofluoromethane	0.687	0.655	4.7	82	0.00
8 T	Diethyl Ether	0.225	0.264	-17.3	113	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.432	0.413	4.4	118	0.00
10 T	Methyl Iodide	0.555	0.587	-5.8	119	0.00
11 T	Tert butyl alcohol	0.091	0.094	-3.3	93	-0.06
12 CM	1,1-Dichloroethene	0.407	0.391	3.9#	116	0.00
13 T	Acrolein	0.084	0.109	-29.8#	140	0.00
14 T	Allyl chloride	0.767	0.701	8.6	98	0.00
15 T	Acrylonitrile	0.241	0.246	-2.1	88	0.00
16 T	Acetone	0.232	0.203	12.5	107	-0.02
17 T	Carbon Disulfide	1.066	0.741	30.5#	83	0.00
18 T	Methyl Acetate	0.735	0.758	-3.1	91	0.00
19 T	Methyl tert-butyl Ether	1.413	1.361	3.7	84	0.00
20 T	Methylene Chloride	0.512	0.457	10.7	83	0.00
21 T	trans-1,2-Dichloroethene	0.476	0.422	11.3	79	0.00
22 T	Diisopropyl ether	1.544	1.448	6.2	82	0.00
23 T	Vinyl Acetate	1.162	1.101	5.2	80	0.00
24 P	1,1-Dichloroethane	0.845	0.795	5.9	83	0.00
25 T	2-Butanone	0.344	0.331	3.8	85	-0.02
26 T	2,2-Dichloropropane	0.605	0.533	11.9	77	0.00
27 T	cis-1,2-Dichloroethene	0.547	0.524	4.2	84	0.00
28 T	Bromochloromethane	0.385	0.324	15.8	75	0.00
29 T	Tetrahydrofuran	0.217	0.217	0.0	87	-0.01
30 C	Chloroform	0.865	0.817	5.5#	83	0.00
31 T	Cyclohexane	0.775	0.661	14.7	73	0.00
32 T	1,1,1-Trichloroethane	0.739	0.683	7.6	80	0.00
33 S	1,2-Dichloroethane-d4	0.526	0.485	7.8	84	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	86	0.00
35 S	Dibromofluoromethane	0.310	0.318	-2.6	91	0.00
36 T	1,1-Dichloropropene	0.453	0.408	9.9	77	0.00
37 T	Ethyl Acetate	0.496	0.448	9.7	84	0.00
38 T	Carbon Tetrachloride	0.469	0.430	8.3	77	0.00
39 T	Methylcyclohexane	0.560	0.495	11.6	76	0.00
40 TM	Benzene	1.321	1.257	4.8	81	0.00
41 T	Methacrylonitrile	0.250	0.265	-6.0	87	0.00
42 TM	1,2-Dichloroethane	0.468	0.437	6.6	80	0.00
43 T	Isopropyl Acetate	0.715	0.712	0.4	84	0.00
44 TM	Trichloroethene	0.369	0.377	-2.2	87	0.00
45 C	1,2-Dichloropropane	0.331	0.328	0.9#	84	0.00
46 T	Dibromomethane	0.231	0.230	0.4	85	0.00
47 T	Bromodichloromethane	0.445	0.437	1.8	81	0.00
48 T	Methyl methacrylate	0.360	0.364	-1.1	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013023\
 Data File : VX033956.D
 Acq On : 30 Jan 2023 19:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 31 04:52:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	86	-0.07
50 S	Toluene-d8	1.131	1.131	0.0	87	0.00
51 T	4-Methyl-2-Pentanone	0.439	0.464	-5.7	86	0.00
52 CM	Toluene	0.853	0.836	2.0#	83	0.00
53 T	t-1,3-Dichloropropene	0.445	0.440	1.1	79	0.00
54 T	cis-1,3-Dichloropropene	0.499	0.502	-0.6	81	0.00
55 T	1,1,2-Trichloroethane	0.338	0.353	-4.4	89	0.00
56 T	Ethyl methacrylate	0.475	0.515	-8.4	86	0.00
57 T	1,3-Dichloropropane	0.547	0.569	-4.0	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.234	0.240	-2.6	85	0.00
59 T	2-Hexanone	0.306	0.342	-11.8	87	0.00
60 T	Dibromochloromethane	0.350	0.373	-6.6	84	0.00
61 T	1,2-Dibromoethane	0.337	0.363	-7.7	84	0.00
62 S	4-Bromofluorobenzene	0.420	0.425	-1.2	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.381	0.383	-0.5	86	0.00
65 PM	Chlorobenzene	1.035	1.053	-1.7	85	0.00
66 T	1,1,1,2-Tetrachloroethane	0.381	0.397	-4.2	85	0.00
67 C	Ethyl Benzene	1.777	1.780	-0.2#	82	0.00
68 T	m/p-Xylenes	0.695	0.710	-2.2	83	0.00
69 T	o-Xylene	0.689	0.707	-2.6	83	0.00
70 T	Styrene	1.121	1.197	-6.8	84	0.00
71 P	Bromoform	0.319	0.332	-4.1	83	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
73 T	Isopropylbenzene	3.645	3.442	5.6	85	0.00
74 T	N-amyl acetate	1.313	1.287	2.0	82	0.00
75 P	1,1,2,2-Tetrachloroethane	1.097	1.074	2.1	89	0.00
76 T	1,2,3-Trichloropropane	0.919	0.862	6.2	89	0.00
77 T	Bromobenzene	0.986	0.954	3.2	87	0.00
78 T	n-propylbenzene	4.080	3.803	6.8	82	0.00
79 T	2-Chlorotoluene	2.483	2.294	7.6	83	0.00
80 T	1,3,5-Trimethylbenzene	3.009	2.842	5.6	84	0.00
81 T	trans-1,4-Dichloro-2-butene	0.296	0.288	2.7	82	0.00
82 T	4-Chlorotoluene	2.818	2.600	7.7	82	0.00
83 T	tert-Butylbenzene	2.987	2.963	0.8	84	0.00
84 T	1,2,4-Trimethylbenzene	2.846	2.852	-0.2	83	0.00
85 T	sec-Butylbenzene	3.573	3.573	0.0	82	0.00
86 T	p-Isopropyltoluene	3.059	3.120	-2.0	83	0.00
87 T	1,3-Dichlorobenzene	1.689	1.707	-1.1	85	0.00
88 T	1,4-Dichlorobenzene	1.721	1.707	0.8	85	0.00
89 T	n-Butylbenzene	2.503	2.487	0.6	81	0.00
90 T	Hexachloroethane	0.503	0.474	5.8	76	0.00
91 T	1,2-Dichlorobenzene	1.630	1.680	-3.1	86	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.204	0.204	0.0	83	0.00
93 T	1,2,4-Trichlorobenzene	1.098	1.194	-8.7	86	0.00
94 T	Hexachlorobutadiene	0.538	0.549	-2.0	85	0.00
95 T	Naphthalene	3.159	3.567	-12.9	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013023\
Data File : VX033956.D
Acq On : 30 Jan 2023 19:56
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 31 04:52:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.099	1.205	-9.6	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6