

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013024\
 Data File : VX040018.D
 Acq On : 30 Jan 2024 14:51
 Operator : JC/MD
 Sample : VX0130WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0130WBL01

Quant Time: Jan 31 05:36:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 31 05:24:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	86529	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	155933	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	150503	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	66328	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65257	46.861	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.720%
35) Dibromofluoromethane	5.385	113	50762	47.821	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.640%
50) Toluene-d8	8.647	98	193689	48.792	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.580%
62) 4-Bromofluorobenzene	11.079	95	74342	48.636	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.280%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013024\
Data File : VX040018.D
Acq On : 30 Jan 2024 14:51
Operator : JC/MD
Sample : VX0130WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0130WBL01

Quant Time: Jan 31 05:36:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 31 05:24:02 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 712
Delta R.T. -0.000 min
Lab File: VX040018.D
Acq: 30 Jan 2024 14:51




Instrument :
MSVOA_X
ClientSampleId :
VX0130WBL01


Tgt Ion:168 Resp: 86529
Ion Ratio Lower Upper
168 100
99 59.6 56.6 85.0



#33
1,2-Dichloroethane-d4
Concen: 46.861 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.000 min
Lab File: VX040018.D
Acq: 30 Jan 2024 14:51

Tgt Ion: 65 Resp: 65257
Ion Ratio Lower Upper
65 100
67 52.4 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX040018.D
Acq: 30 Jan 2024 14:51




Instrument :
MSVOA_X
ClientSampleId :
VX0130WBL01


Tgt Ion:114 Resp: 155933
Ion Ratio Lower Upper
114 100
63 21.2 0.0 45.0
88 15.2 0.0 36.6



#35
Dibromofluoromethane
Concen: 47.821 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX040018.D
Acq: 30 Jan 2024 14:51

Tgt Ion:113 Resp: 50762
Ion Ratio Lower Upper
113 100
111 102.7 82.9 124.3
192 16.5 13.3 19.9









#50
Toluene-d8
Concen: 48.792 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX040018.D
Acq: 30 Jan 2024 14:51




Instrument :
MSVOA_X
ClientSampleId :
VX0130WBL01


Tgt Ion: 98 Resp: 193689
Ion Ratio Lower Upper
98 100
100 64.2 53.1 79.7



#62
4-Bromofluorobenzene
Concen: 48.636 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX040018.D
Acq: 30 Jan 2024 14:51

Tgt Ion: 95 Resp: 74342
Ion Ratio Lower Upper
95 100
174 69.0 0.0 135.6
176 66.7 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX040018.D
Acq: 30 Jan 2024 14:51

Instrument :
MSVOA_X
ClientSampleId :
VX0130WBL01

Tgt Ion:117 Resp: 150503
Ion Ratio Lower Upper
117 100
82 58.2 46.2 69.4
119 33.2 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.006 min
Lab File: VX040018.D
Acq: 30 Jan 2024 14:51

Tgt Ion:152 Resp: 66328
Ion Ratio Lower Upper
152 100
115 67.2 43.3 129.9
150 161.2 0.0 345.0

