



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013024\
Data File : VX040008.D
Acq On : 30 Jan 2024 08:50
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Title : SW846 8260
Last Update : Wed Jan 31 07:12:04 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.2	5159	PASS
75	95	30	60	54.5	13896	PASS
95	95	100	100	100.0	25515	PASS
96	95	5	9	6.4	1623	PASS
173	174	0.00	2	0.8	138	PASS
174	95	50	100	68.4	17440	PASS
175	174	5	9	7.2	1256	PASS
176	174	95	101	96.8	16876	PASS
177	176	5	9	6.9	1162	PASS