

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013025\
 Data File : VX044780.D
 Acq On : 30 Jan 2025 13:55
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jan 31 00:42:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011625W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Jan 17 01:21:41 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.898	128	24187	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.757	114	130445	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	112829	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	63658	27.263	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	90.867%#
60) 4-Bromofluorobenzene	11.079	95	61462	28.045	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	93.500%
63) Toluene-d8	8.647	98	186168	30.337	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.133%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013025\
Data File : VX044780.D
Acq On : 30 Jan 2025 13:55
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jan 31 00:42:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011625W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Jan 17 01:21:41 2025
Response via : Initial Calibration

