

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013122\
 Data File : VX026681.D
 Acq On : 31 Jan 2022 15:31
 Operator : JC/MD
 Sample : PB142369ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142369ZHE#04

Quant Time: Jan 31 16:33:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	135026	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	226114	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208186	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90108	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85408	45.280	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.560%
35) Dibromofluoromethane	5.391	113	69688	47.726	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.460%
50) Toluene-d8	8.653	98	269738	51.263	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.520%
62) 4-Bromofluorobenzene	11.079	95	98139	52.503	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.000%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	11846	12.315	ug/l	98
43) Isopropyl Acetate	6.342	43	11750	3.120	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013122\
Data File : VX026681.D
Acq On : 31 Jan 2022 15:31
Operator : JC/MD
Sample : PB142369ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142369ZHE#04

Quant Time: Jan 31 16:33:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

