

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013122\
 Data File : VX026683.D
 Acq On : 31 Jan 2022 16:17
 Operator : JC/MD
 Sample : PB142369ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142369ZHE#06

Quant Time: Feb 01 06:28:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	138467	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	233589	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208773	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86711	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88775	45.896	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.800%
35) Dibromofluoromethane	5.391	113	72421	48.010	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.020%
50) Toluene-d8	8.653	98	275109	50.611	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.220%
62) 4-Bromofluorobenzene	11.079	95	96458	49.952	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.900%
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	11370	11.527	ug/l	97
43) Isopropyl Acetate	6.348	43	11951	3.071	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013122\
Data File : VX026683.D
Acq On : 31 Jan 2022 16:17
Operator : JC/MD
Sample : PB142369ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142369ZHE#06

Quant Time: Feb 01 06:28:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

