

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013123\
 Data File : VX033967.D
 Acq On : 31 Jan 2023 13:32
 Operator : JC/MD
 Sample : 01358-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DR-2-012723

Quant Time: Feb 01 02:11:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	159238	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	230105	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206913	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99686	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	76215	45.457	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.920%
35) Dibromofluoromethane	5.385	113	69642	48.741	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.480%
50) Toluene-d8	8.647	98	260688	50.075	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.140%
62) 4-Bromofluorobenzene	11.079	95	91859	47.524	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.040%
Target Compounds						
						Qvalue
16) Acetone	2.373	43	13832	10.850	ug/l	96
37) Ethyl Acetate	4.714	43	21922	8.654	ug/l	99
43) Isopropyl Acetate	6.336	43	90375	27.452	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013123\
 Data File : VX033967.D
 Acq On : 31 Jan 2023 13:32
 Operator : JC/MD
 Sample : 01358-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DR-2-012723

Quant Time: Feb 01 02:11:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

