

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013123\
 Data File : VX033969.D
 Acq On : 31 Jan 2023 14:18
 Operator : JC/MD
 Sample : 01358-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JC-3-012723

Quant Time: Feb 01 02:12:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	165296	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	239421	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	211297	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100517	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	79145	45.475	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.940%
35) Dibromofluoromethane	5.385	113	72697	48.899	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.800%
50) Toluene-d8	8.647	98	271720	50.163	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.320%
62) 4-Bromofluorobenzene	11.079	95	93946	46.713	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.420%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	13926	10.131	ug/l	98
37) Ethyl Acetate	4.715	43	20937	7.695	ug/l	97
43) Isopropyl Acetate	6.336	43	88189	25.746	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013123\
Data File : VX033969.D
Acq On : 31 Jan 2023 14:18
Operator : JC/MD
Sample : 01358-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JC-3-012723

Quant Time: Feb 01 02:12:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

