

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013123\
 Data File : VX033971.D
 Acq On : 31 Jan 2023 15:04
 Operator : JC/MD
 Sample : 01360-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-3-WC

Quant Time: Feb 01 02:13:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	165272	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	239865	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213433	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101726	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	79367	45.609	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.220%
35) Dibromofluoromethane	5.385	113	72862	48.919	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.840%
50) Toluene-d8	8.647	98	273356	50.371	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.740%
62) 4-Bromofluorobenzene	11.079	95	95017	47.158	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.320%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	61825	89.825	ug/l	93
37) Ethyl Acetate	4.709	43	21648	8.039	ug/l	98
43) Isopropyl Acetate	6.336	43	93071	27.121	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013123\
Data File : VX033971.D
Acq On : 31 Jan 2023 15:04
Operator : JC/MD
Sample : 01360-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-3-WC

Quant Time: Feb 01 02:13:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

