

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013123\
 Data File : VX033974.D
 Acq On : 31 Jan 2023 16:25
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Feb 01 02:15:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	165882	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	235649	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213868	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	106960	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	78003	44.660	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.320%
35) Dibromofluoromethane	5.379	113	70800	48.385	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.780%
50) Toluene-d8	8.647	98	265849	49.865	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.720%
62) 4-Bromofluorobenzene	11.079	95	96046	48.522	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.040%
Target Compounds						
					Qvalue	
40) Benzene	6.037	78	22504	3.615	ug/l	98
52) Toluene	8.720	92	15951	3.968	ug/l	99
67) Ethyl Benzene	10.195	91	13419	1.765	ug/l	97
68) m/p-Xylenes	10.299	106	6327	2.128	ug/l	97
95) Naphthalene	13.774	128	44667	6.609	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013123\
Data File : VX033974.D
Acq On : 31 Jan 2023 16:25
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Feb 01 02:15:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

