

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013123\
 Data File : VX033958.D
 Acq On : 31 Jan 2023 09:00
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 01 02:03:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.395	0.342	13.4	80	0.00
3 P	Chloromethane	0.443	0.409	7.7	78	0.00
4 C	Vinyl Chloride	0.424	0.422	0.5#	82	0.00
5 T	Bromomethane	0.243	0.210	13.6	86	0.00
6 T	Chloroethane	0.262	0.223	14.9	78	0.00
7 T	Trichlorofluoromethane	0.687	0.630	8.3	85	0.00
8 T	Diethyl Ether	0.225	0.254	-12.9	117	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.432	0.421	2.5	130	0.00
10 T	Methyl Iodide	0.555	0.586	-5.6	128	0.00
11 T	Tert butyl alcohol	0.091	0.081	11.0	86	-0.06
12 CM	1,1-Dichloroethene	0.407	0.396	2.7#	127	0.00
13 T	Acrolein	0.084	0.098	-16.7	136	0.00
14 T	Allyl chloride	0.767	0.682	11.1	103	0.00
15 T	Acrylonitrile	0.241	0.241	0.0	93	-0.01
16 T	Acetone	0.232	0.201	13.4	114	-0.02
17 T	Carbon Disulfide	1.066	0.794	25.5#	96	0.00
18 T	Methyl Acetate	0.735	0.733	0.3	95	-0.01
19 T	Methyl tert-butyl Ether	1.413	1.347	4.7	90	0.00
20 T	Methylene Chloride	0.512	0.457	10.7	90	0.00
21 T	trans-1,2-Dichloroethene	0.476	0.427	10.3	87	0.00
22 T	Diisopropyl ether	1.544	1.475	4.5	90	0.00
23 T	Vinyl Acetate	1.162	1.090	6.2	86	0.00
24 P	1,1-Dichloroethane	0.845	0.789	6.6	89	0.00
25 T	2-Butanone	0.344	0.317	7.8	88	-0.02
26 T	2,2-Dichloropropane	0.605	0.457	24.5	71	0.00
27 T	cis-1,2-Dichloroethene	0.547	0.529	3.3	91	0.00
28 T	Bromochloromethane	0.385	0.353	8.3	89	0.00
29 T	Tetrahydrofuran	0.217	0.209	3.7	90	-0.01
30 C	Chloroform	0.865	0.825	4.6#	91	0.00
31 T	Cyclohexane	0.775	0.668	13.8	79	0.00
32 T	1,1,1-Trichloroethane	0.739	0.695	6.0	88	0.00
33 S	1,2-Dichloroethane-d4	0.526	0.477	9.3	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.310	0.313	-1.0	97	0.00
36 T	1,1-Dichloropropene	0.453	0.409	9.7	84	0.00
37 T	Ethyl Acetate	0.496	0.436	12.1	88	-0.01
38 T	Carbon Tetrachloride	0.469	0.443	5.5	86	0.00
39 T	Methylcyclohexane	0.560	0.491	12.3	81	0.00
40 TM	Benzene	1.321	1.253	5.1	88	0.00
41 T	Methacrylonitrile	0.250	0.257	-2.8	91	0.00
42 TM	1,2-Dichloroethane	0.468	0.436	6.8	86	0.00
43 T	Isopropyl Acetate	0.715	0.691	3.4	88	0.00
44 TM	Trichloroethene	0.369	0.374	-1.4	93	0.00
45 C	1,2-Dichloropropane	0.331	0.323	2.4#	90	0.00
46 T	Dibromomethane	0.231	0.226	2.2	91	0.00
47 T	Bromodichloromethane	0.445	0.449	-0.9	91	0.00
48 T	Methyl methacrylate	0.360	0.356	1.1	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013123\
 Data File : VX033958.D
 Acq On : 31 Jan 2023 09:00
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 01 02:03:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	81	-0.06
50 S	Toluene-d8	1.131	1.096	3.1	92	0.00
51 T	4-Methyl-2-Pentanone	0.439	0.450	-2.5	91	0.00
52 CM	Toluene	0.853	0.822	3.6#	89	0.00
53 T	t-1,3-Dichloropropene	0.445	0.440	1.1	86	0.00
54 T	cis-1,3-Dichloropropene	0.499	0.491	1.6	86	0.00
55 T	1,1,2-Trichloroethane	0.338	0.343	-1.5	94	0.00
56 T	Ethyl methacrylate	0.475	0.506	-6.5	92	0.00
57 T	1,3-Dichloropropane	0.547	0.568	-3.8	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.234	0.237	-1.3	91	0.00
59 T	2-Hexanone	0.306	0.322	-5.2	89	0.00
60 T	Dibromochloromethane	0.350	0.385	-10.0	94	0.00
61 T	1,2-Dibromoethane	0.337	0.361	-7.1	91	0.00
62 S	4-Bromofluorobenzene	0.420	0.419	0.2	94	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.381	0.384	-0.8	93	0.00
65 PM	Chlorobenzene	1.035	1.043	-0.8	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.381	0.409	-7.3	95	0.00
67 C	Ethyl Benzene	1.777	1.788	-0.6#	89	0.00
68 T	m/p-Xylenes	0.695	0.712	-2.4	90	0.00
69 T	o-Xylene	0.689	0.696	-1.0	89	0.00
70 T	Styrene	1.121	1.178	-5.1	90	0.00
71 P	Bromoform	0.319	0.344	-7.8	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.645	3.358	7.9	90	0.00
74 T	N-amyl acetate	1.313	1.224	6.8	85	0.00
75 P	1,1,2,2-Tetrachloroethane	1.097	1.034	5.7	93	0.00
76 T	1,2,3-Trichloropropane	0.919	0.929	-1.1	104	0.00
77 T	Bromobenzene	0.986	0.923	6.4	91	0.00
78 T	n-propylbenzene	4.080	3.739	8.4	87	0.00
79 T	2-Chlorotoluene	2.483	2.232	10.1	88	0.00
80 T	1,3,5-Trimethylbenzene	3.009	2.817	6.4	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.296	0.268	9.5	83	0.00
82 T	4-Chlorotoluene	2.818	2.548	9.6	88	0.00
83 T	tert-Butylbenzene	2.987	2.905	2.7	89	0.00
84 T	1,2,4-Trimethylbenzene	2.846	2.788	2.0	88	0.00
85 T	sec-Butylbenzene	3.573	3.525	1.3	88	0.00
86 T	p-Isopropyltoluene	3.059	3.048	0.4	88	0.00
87 T	1,3-Dichlorobenzene	1.689	1.676	0.8	91	0.00
88 T	1,4-Dichlorobenzene	1.721	1.688	1.9	91	0.00
89 T	n-Butylbenzene	2.503	2.446	2.3	86	0.00
90 T	Hexachloroethane	0.503	0.501	0.4	87	0.00
91 T	1,2-Dichlorobenzene	1.630	1.665	-2.1	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.204	0.205	-0.5	91	0.00
93 T	1,2,4-Trichlorobenzene	1.098	1.151	-4.8	91	0.00
94 T	Hexachlorobutadiene	0.538	0.533	0.9	90	0.00
95 T	Naphthalene	3.159	3.359	-6.3	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013123\
Data File : VX033958.D
Acq On : 31 Jan 2023 09:00
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 01 02:03:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.099	1.167	-6.2	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6