

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013124\
 Data File : VX040039.D
 Acq On : 31 Jan 2024 11:41
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Feb 01 04:21:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 31 07:12:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	89348	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	158793	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	144489	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69478	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67323	46.819	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.640%
35) Dibromofluoromethane	5.385	113	53141	49.161	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.320%
50) Toluene-d8	8.647	98	192045	47.507	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.020%
62) 4-Bromofluorobenzene	11.079	95	72856	46.806	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013124\
Data File : VX040039.D
Acq On : 31 Jan 2024 11:41
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Feb 01 04:21:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 31 07:12:04 2024
Response via : Initial Calibration

