

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013124\
 Data File : VX040048.D
 Acq On : 31 Jan 2024 15:06
 Operator : JC/MD
 Sample : P1282-03 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
ClientSampleId :
 1242

Quant Time: Feb 01 04:23:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 31 07:12:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.556	168	90386	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	159439	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	150601	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74069	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66336	45.603	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.200%
35) Dibromofluoromethane	5.385	113	53167	48.986	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.980%
50) Toluene-d8	8.647	98	195369	48.133	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.260%
62) 4-Bromofluorobenzene	11.079	95	70950	45.396	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013124\
 Data File : VX040048.D
 Acq On : 31 Jan 2024 15:06
 Operator : JC/MD
 Sample : P1282-03 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1242

Quant Time: Feb 01 04:23:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 31 07:12:04 2024
 Response via : Initial Calibration

