



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013124\
Data File : VX040033.D
Acq On : 31 Jan 2024 08:31
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Title : SW846 8260
Last Update : Thu Feb 01 05:44:39 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.1	6041	PASS
75	95	30	60	55.1	15063	PASS
95	95	100	100	100.0	27328	PASS
96	95	5	9	6.9	1893	PASS
173	174	0.00	2	0.8	167	PASS
174	95	50	100	72.5	19811	PASS
175	174	5	9	6.9	1376	PASS
176	174	95	101	95.8	18973	PASS
177	176	5	9	6.3	1187	PASS