

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013125\
 Data File : VX044795.D
 Acq On : 31 Jan 2025 14:40
 Operator : JC/MD
 Sample : Q1189-03ME
 Misc : 15.32g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Feb 03 07:34:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML012825WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Feb 03 07:31:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	304183	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	274837	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	133231	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	65851	28.760	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	57.520%#
7) Chloroethane-d5	1.648	69	13843	15.914	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	31.820%#
11) 1,1-Dichloroethene-d2	2.282	65	44985	45.963	ug/L	-0.01
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.920%
21) 2-Butanone-d5	4.477	46	119665	122.249	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	122.250%
24) Chloroform-d	5.056	84	184522	45.861	ug/L	0.01
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.720%
26) 1,2-Dichloroethane-d4	5.952	65	115023	47.627	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.260%
32) Benzene-d6	5.964	84	404927	47.120	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.240%
36) 1,2-Dichloropropane-d6	7.306	67	127417	46.844	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.680%
41) Toluene-d8	8.647	98	360440	47.390	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.780%
43) trans-1,3-Dichloroprop...	8.952	79	56644	43.902	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.800%
47) 2-Hexanone-d5	9.391	63	80685	100.229	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.230%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	167629	49.593	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.180%
66) 1,2-Dichlorobenzene-d4	12.317	152	123510	47.378	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013125\
Data File : VX044795.D
Acq On : 31 Jan 2025 14:40
Operator : JC/MD
Sample : Q1189-03ME
Misc : 15.32g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Feb 03 07:34:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML012825WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Feb 03 07:31:09 2025
Response via : Initial Calibration

