

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013125\
 Data File : VX044784.D
 Acq On : 31 Jan 2025 10:08
 Operator : JC/MD
 Sample : VX0131MBL01
 Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK757

Quant Time: Feb 03 07:31:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML012825WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Feb 03 07:31:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.751	114	262435	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.049	117	231290	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	99094	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	82823	41.926	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	83.860%	
7) Chloroethane-d5	1.648	69	49243	65.616	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	131.240%#	
11) 1,1-Dichloroethene-d2	2.300	65	38912	46.082	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	92.160%	
21) 2-Butanone-d5	4.453	46	100587	119.106	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	119.110%	
24) Chloroform-d	5.044	84	149151	42.967	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	85.940%	
26) 1,2-Dichloroethane-d4	5.946	65	94739	45.469	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	90.940%	
32) Benzene-d6	5.964	84	316275	43.733	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	87.460%	
36) 1,2-Dichloropropane-d6	7.300	67	101980	44.551	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	89.100%	
41) Toluene-d8	8.641	98	277236	43.313	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	86.620%	
43) trans-1,3-Dichloroprop...	8.946	79	45870	42.245	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	84.500%	
47) 2-Hexanone-d5	9.378	63	64000	94.471	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	94.470%	
56) 1,1,2,2-Tetrachloroeth...	11.189	84	134104	47.145	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	94.280%	
66) 1,2-Dichlorobenzene-d4	12.317	152	90861	46.861	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	93.720%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX013125\
Data File : VX044784.D
Acq On : 31 Jan 2025 10:08
Operator : JC/MD
Sample : VX0131MBL01
Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK757

Quant Time: Feb 03 07:31:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML012825WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Feb 03 07:31:09 2025
Response via : Initial Calibration

