

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020119\
 Data File : VX007342.D
 Acq On : 01 Feb 2019 17:46
 Operator : JC/SP
 Sample : K1314-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Feb 02 04:00:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	508689	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	791062	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	749980	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	360166	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	274410	50.82	ug/l	0.00
Spiked Amount						
			Recovery	=	101.64%	
35) Dibromofluoromethane	5.51	113	248309	48.24	ug/l	0.00
Spiked Amount						
			Recovery	=	96.48%	
50) Toluene-d8	8.71	98	948830	49.69	ug/l	0.00
Spiked Amount						
			Recovery	=	99.38%	
62) 4-Bromofluorobenzene	11.13	95	339630	48.71	ug/l	0.00
Spiked Amount						
			Recovery	=	97.42%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020119\
Data File : VX007342.D
Acq On : 01 Feb 2019 17:46
Operator : JC/SP
Sample : K1314-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Feb 02 04:00:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

