

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020119\
 Data File : VX007336.D
 Acq On : 01 Feb 2019 15:07
 Operator : JC/SP
 Sample : K1296-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-TB01-20190129

Quant Time: Feb 02 03:39:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	522849	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	795320	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	762129	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	373473	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	276527	49.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.66%	
35) Dibromofluoromethane	5.50	113	255058	49.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.58%	
50) Toluene-d8	8.71	98	960790	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
62) 4-Bromofluorobenzene	11.13	95	348302	49.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.38%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020119\
Data File : VX007336.D
Acq On : 01 Feb 2019 15:07
Operator : JC/SP
Sample : K1296-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-TB01-20190129

Quant Time: Feb 02 03:39:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.01 min

Lab File: VX007336.D

Acq: 01 Feb 2019 15:07

Instrument :

MSVOA_X

ClientSampleId :


BP-TT-TB01-20190129

Tot Ion:168 Resp: 522849

Ion Ratio Lower Upper

168 100

99 49.8 37.9 56.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

200000



150000

100000

50000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 49.828 ug/l

RT: 6.07 min Scan# 806

Delta R.T. 0.00 min

Lab File: VX007336.D


Acq: 01 Feb 2019 15:07

Tgt Ion: 65 Resp: 276527

Ion Ratio Lower Upper

65 100

67 53.4 0.0 108.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07

120000

100000

80000


60000


40000

20000

0

Time-->









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. 0.00 min
 Lab File: VX007336.D
 Acq: 01 Feb 2019 15:07




Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-TB01-20190129


Tot Ion	Ratio	Lower	Upper
114	100		
63	18.2	0.0	35.6
88	13.8	0.0	29.6



#35
 Dibromofluoromethane
 Concen: 49.288 ug/l
 RT: 5.50 min Scan# 713
 Delta R.T. -0.01 min
 Lab File: VX007336.D
 Acq: 01 Feb 2019 15:07

Tgt Ion	Ratio	Lower	Upper
113	100		
111	100.0	81.8	122.8
192	22.1	18.5	27.7





#50

Toluene-d8

Concen: 50.042 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX007336.D

Acq: 01 Feb 2019 15:07

Instrument :

MSVOA_X

ClientSampleId :


BP-TT-TB01-20190129

Tot Ion: 98 Resp: 960790

Ion Ratio Lower Upper




98 100

100 63.8 51.6 77.4



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX



#62

4-Bromofluorobenzene

Concen: 49.688 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.00 min

Lab File: VX007336.D

Acq: 01 Feb 2019 15:07


Tgt Ion: 95 Resp: 348302

Ion Ratio Lower Upper

95 100

174 93.2 0.0 189.2



176 89.9 0.0 186.2




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. 0.00 min
Lab File: VX007336.D
Acq: 01 Feb 2019 15:07

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-TB01-20190129

Tot Ion:117 Resp: 762129
Ion Ratio Lower Upper
117 100
82 51.7 42.1 63.1
119 31.5 26.5 39.7




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX007336.D
Acq: 01 Feb 2019 15:07

Tgt Ion:152 Resp: 373473
Ion Ratio Lower Upper
152 100
115 54.6 38.0 114.0
150 157.2 0.0 346.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

