

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020122\
 Data File : VX026704.D
 Acq On : 01 Feb 2022 14:11
 Operator : JC/MD
 Sample : N1325-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TEX-3

Quant Time: Feb 02 04:35:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	132536	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	222830	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	199334	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84046	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84734	45.767	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.540%
35) Dibromofluoromethane	5.391	113	68835	47.836	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.680%
50) Toluene-d8	8.653	98	260679	50.272	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.540%
62) 4-Bromofluorobenzene	11.085	95	92834	50.397	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.800%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	10852	11.494	ug/l	97
20) Methylene Chloride	2.794	84	1749	1.121	ug/l	96
43) Isopropyl Acetate	6.342	43	11509	3.101	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020122\
Data File : VX026704.D
Acq On : 01 Feb 2022 14:11
Operator : JC/MD
Sample : N1325-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TEX-3

Quant Time: Feb 02 04:35:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

