

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020122\
 Data File : VX026722.D
 Acq On : 01 Feb 2022 21:06
 Operator : JC/MD
 Sample : PB142403ZHE#15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142403ZHE#15

Quant Time: Feb 02 04:38:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	128268	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	216625	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	196720	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84764	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81984	45.755	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.500%
35) Dibromofluoromethane	5.391	113	67171	48.017	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.040%
50) Toluene-d8	8.653	98	255917	50.767	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.540%
62) 4-Bromofluorobenzene	11.079	95	92954	51.908	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.820%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	8115	8.881	ug/l	99
18) Methyl Acetate	2.709	43	2137	1.176	ug/l #	83
20) Methylene Chloride	2.794	84	2872	1.902	ug/l	86
43) Isopropyl Acetate	6.342	43	19459	5.392	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020122\
Data File : VX026722.D
Acq On : 01 Feb 2022 21:06
Operator : JC/MD
Sample : PB142403ZHE#15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142403ZHE#15

Quant Time: Feb 02 04:38:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

