

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020124\
 Data File : VX040071.D
 Acq On : 01 Feb 2024 15:10
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Feb 02 01:41:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	84778	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	150494	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	149975	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72120	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62491	45.801	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.600%
35) Dibromofluoromethane	5.385	113	49306	48.129	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.260%
50) Toluene-d8	8.646	98	181145	47.281	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.560%
62) 4-Bromofluorobenzene	11.079	95	72473	49.127	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020124\
Data File : VX040071.D
Acq On : 01 Feb 2024 15:10
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Feb 02 01:41:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

