

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020222\
 Data File : VX026735.D
 Acq On : 02 Feb 2022 14:37
 Operator : JC/MD
 Sample : VX0202WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0202WBSD03

Quant Time: Feb 03 03:06:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	92843	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	149714	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128688	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59140	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	56737	43.747	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.500%
35) Dibromofluoromethane	5.385	113	45513	47.076	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.160%
50) Toluene-d8	8.647	98	165657	47.549	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.100%
62) 4-Bromofluorobenzene	11.079	95	59445	48.031	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.060%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.172	85	15546	14.180	ug/l	98
3) Chloromethane	1.294	50	15202	15.126	ug/l	100
4) Vinyl Chloride	1.374	62	16207	15.359	ug/l	98
5) Bromomethane	1.617	94	7709	15.837	ug/l	98
6) Chloroethane	1.691	64	9985	18.713	ug/l	99
7) Trichlorofluoromethane	1.892	101	26116	15.676	ug/l	97
8) Diethyl Ether	2.136	74	9927	16.839	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.331	101	18419	18.143	ug/l	99
10) Methyl Iodide	2.459	142	21266	22.173	ug/l	96
11) Tert butyl alcohol	2.983	59	15931	88.300	ug/l #	94
12) 1,1-Dichloroethene	2.325	96	16467	16.983	ug/l	91
13) Acrolein	2.239	56	18372	95.830	ug/l	100
14) Allyl chloride	2.666	41	30418	17.059	ug/l	97
15) Acrylonitrile	3.068	53	53993	97.217	ug/l	98
16) Acetone	2.386	43	47275	71.478	ug/l	97
17) Carbon Disulfide	2.514	76	38194	13.492	ug/l	100
18) Methyl Acetate	2.709	43	25701	19.546	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	63281	17.986	ug/l	98
20) Methylene Chloride	2.794	84	19341	17.699	ug/l	97
21) trans-1,2-Dichloroethene	3.099	96	17885	17.039	ug/l	96
22) Diisopropyl ether	3.763	45	66827	18.450	ug/l	97
23) Vinyl Acetate	3.721	43	274561	92.088	ug/l	98
24) 1,1-Dichloroethane	3.617	63	34351	17.561	ug/l	100
25) 2-Butanone	4.562	43	74110	86.162	ug/l	99
26) 2,2-Dichloropropane	4.477	77	25253	17.416	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	22496	18.778	ug/l	96
28) Bromochloromethane	4.897	49	14436	17.973	ug/l	100
29) Tetrahydrofuran	5.007	42	48972	95.865	ug/l	100
30) Chloroform	5.099	83	37014	18.211	ug/l	100
31) Cyclohexane	5.470	56	30067	15.718	ug/l	98
32) 1,1,1-Trichloroethane	5.385	97	31240	17.474	ug/l	99
36) 1,1-Dichloropropene	5.696	75	25768	16.883	ug/l	97
37) Ethyl Acetate	4.715	43	28836	18.677	ug/l	98
38) Carbon Tetrachloride	5.678	117	27073	17.245	ug/l	99
39) Methylcyclohexane	7.385	83	31553	17.139	ug/l	99
40) Benzene	6.044	78	78140	18.523	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020222\
 Data File : VX026735.D
 Acq On : 02 Feb 2022 14:37
 Operator : JC/MD
 Sample : VX0202WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0202WBSD03

Quant Time: Feb 03 03:06:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	16839	19.254	ug/l	97
42) 1,2-Dichloroethane	6.092	62	29163	17.669	ug/l	98
43) Isopropyl Acetate	6.342	43	47012	18.850	ug/l	99
44) Trichloroethene	7.129	130	23146	19.336	ug/l	99
45) 1,2-Dichloropropane	7.434	63	20383	19.357	ug/l	95
46) Dibromomethane	7.586	93	14005	18.867	ug/l	98
47) Bromodichloromethane	7.824	83	27742	18.497	ug/l	96
48) Methyl methacrylate	7.696	41	22749	18.406	ug/l	96
49) 1,4-Dioxane	7.683	88	7103	350.198	ug/l	99
51) 4-Methyl-2-Pentanone	8.574	43	147479	99.699	ug/l	98
52) Toluene	8.720	92	49276	19.048	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	26649	17.876	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	30907	19.186	ug/l	96
55) 1,1,2-Trichloroethane	9.153	97	20464	20.271	ug/l	97
56) Ethyl methacrylate	9.116	69	31510	20.367	ug/l	94
57) 1,3-Dichloropropane	9.311	76	34638	19.791	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	73510	90.614	ug/l	98
59) 2-Hexanone	9.433	43	106572	94.130	ug/l	98
60) Dibromochloromethane	9.525	129	20640	19.620	ug/l	97
61) 1,2-Dibromoethane	9.610	107	21047	20.008	ug/l	98
64) Tetrachloroethene	9.275	164	25061	20.615	ug/l	94
65) Chlorobenzene	10.079	112	52607	19.412	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.165	131	19674	19.538	ug/l	100
67) Ethyl Benzene	10.195	91	92623	18.808	ug/l	97
68) m/p-Xylenes	10.305	106	70884	38.462	ug/l	97
69) o-Xylene	10.640	106	34597	19.455	ug/l	97
70) Styrene	10.659	104	57492	19.813	ug/l	98
71) Bromoform	10.799	173	13265	18.685	ug/l #	99
73) Isopropylbenzene	10.963	105	91629	18.902	ug/l	99
74) N-amyl acetate	10.841	43	35666	19.058	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.213	83	25160	19.337	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	29925	22.982	ug/l	94
77) Bromobenzene	11.201	156	21533	19.389	ug/l	95
78) n-propylbenzene	11.305	91	103487	18.720	ug/l	100
79) 2-Chlorotoluene	11.366	91	61859	18.453	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	75826	18.973	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	6815	19.841	ug/l	99
82) 4-Chlorotoluene	11.457	91	71223	18.487	ug/l	98
83) tert-Butylbenzene	11.713	119	70778	18.971	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	75273	18.831	ug/l	98
85) sec-Butylbenzene	11.890	105	92859	19.242	ug/l	99
86) p-Isopropyltoluene	12.012	119	77804	19.217	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	39543	18.996	ug/l	99
88) 1,4-Dichlorobenzene	12.042	146	39355	18.679	ug/l	98
89) n-Butylbenzene	12.335	91	65764	18.928	ug/l	99
90) Hexachloroethane	12.542	117	11145	17.954	ug/l	94
91) 1,2-Dichlorobenzene	12.335	146	38808	19.728	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	5485	17.442	ug/l	94
93) 1,2,4-Trichlorobenzene	13.591	180	24522	19.443	ug/l	99
94) Hexachlorobutadiene	13.725	225	10673	20.047	ug/l	97
95) Naphthalene	13.780	128	83436	19.834	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	24920	20.091	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020222\
Data File : VX026735.D
Acq On : 02 Feb 2022 14:37
Operator : JC/MD
Sample : VX0202WBSD02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Quant Time: Feb 03 03:06:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Feb 03 03:06:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03







Tgt Ion:168 Resp: 92843
Ion Ratio Lower Upper
168 100
99 59.5 50.3 75.5



#2
Dichlorodifluoromethane
Concen: 14.180 ug/l
RT: 1.172 min Scan# 14
Delta R.T. -0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion: 85 Resp: 15546
Ion Ratio Lower Upper
85 100
87 31.3 16.4 49.1









#3
Chloromethane
Concen: 15.126 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Tgt Ion: 50 Resp: 15202
Ion Ratio Lower Upper
50 100
52 32.9 26.1 39.1



#4
Vinyl Chloride
Concen: 15.359 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.006 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion: 62 Resp: 16207
Ion Ratio Lower Upper
62 100
64 31.7 26.2 39.2









#5
Bromomethane
Concen: 15.837 ug/l
RT: 1.617 min Scan# 81
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Tgt Ion: 94 Resp: 7709
Ion Ratio Lower Upper
94 100
96 93.1 75.9 113.9



#6
Chloroethane
Concen: 18.713 ug/l
RT: 1.691 min Scan# 99
Delta R.T. -0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion: 64 Resp: 9985
Ion Ratio Lower Upper
64 100
66 31.4 25.4 38.2









#7
 Trichlorofluoromethane
 Concen: 15.676 ug/l
 RT: 1.892 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX026735.D
 Acq: 02 Feb 2022 14:37




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0202WBSD03


Tgt Ion:101 Resp: 26116
 Ion Ratio Lower Upper
 101 100
 103 66.2 51.3 76.9



#8
 Diethyl Ether
 Concen: 16.839 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. 0.000 min
 Lab File: VX026735.D
 Acq: 02 Feb 2022 14:37

Tgt Ion: 74 Resp: 9927
 Ion Ratio Lower Upper
 74 100
 45 101.9 50.7 152.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.143 ug/l
RT: 2.331 min Scan# 205
Delta R.T. -0.006 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Tgt Ion:101 Resp: 18419
Ion Ratio Lower Upper
101 100
85 42.5 35.3 52.9
151 76.9 61.4 92.2



#10
Methyl Iodide
Concen: 22.173 ug/l
RT: 2.459 min Scan# 225
Delta R.T. -0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion:142 Resp: 21266
Ion Ratio Lower Upper
142 100
127 44.4 38.1 57.1
141 13.9 11.2 16.8









#11
Tert butyl alcohol
Concen: 88.300 ug/l
RT: 2.983 min Scan# 311
Delta R.T. -0.024 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03

Tgt Ion: 59 Resp: 15931
Ion Ratio Lower Upper
59 100
57 11.0 7.2 10.8#



#12
1,1-Dichloroethene
Concen: 16.983 ug/l
RT: 2.325 min Scan# 203
Delta R.T. -0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion: 96 Resp: 16467
Ion Ratio Lower Upper
96 100
61 156.9 138.5 207.7
98 65.4 51.8 77.6





#13

Acrolein

Concen: 95.830 ug/l

RT: 2.239 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37

Instrument :

MSVOA_X

ClientSampleId :





VX0202WBSD03

Tgt Ion: 56 Resp: 18372

Ion Ratio Lower Upper

56 100

55 71.3 56.9 85.3



#14

Allyl chloride

Concen: 17.059 ug/l

RT: 2.666 min Scan# 259

Delta R.T. -0.006 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37




Tgt Ion: 41 Resp: 30418


Ion Ratio Lower Upper

41 100

39 63.7 54.2 81.4

76 31.5 25.2 37.8









#15
Acrylonitrile
Concen: 97.217 ug/l
RT: 3.068 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Tgt Ion:	53	Resp:	53993
Ion	Ratio	Lower	Upper
53	100		
52	82.0	66.6	100.0
51	35.9	29.5	44.3



#16
Acetone
Concen: 71.478 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.006 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion:	43	Resp:	47275
Ion	Ratio	Lower	Upper
43	100		
58	30.2	23.0	34.6









#17
Carbon Disulfide
Concen: 13.492 ug/l
RT: 2.514 min Scan# 21
Delta R.T. -0.006 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03

Tgt Ion: 76 Resp: 38194
Ion Ratio Lower Upper
76 100
78 9.3 7.5 11.3



#18
Methyl Acetate
Concen: 19.546 ug/l
RT: 2.709 min Scan# 266
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion: 43 Resp: 25701
Ion Ratio Lower Upper
43 100
74 22.9 18.3 27.5









#19
Methyl tert-butyl Ether
Concen: 17.986 ug/l
RT: 3.111 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03

Tgt Ion: 73 Resp: 63281
Ion Ratio Lower Upper
73 100
57 21.0 17.4 26.2



#20
Methylene Chloride
Concen: 17.699 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion: 84 Resp: 19341
Ion Ratio Lower Upper
84 100
49 122.8 101.5 152.3
51 37.5 31.2 46.8
86 63.1 51.3 76.9









#21
trans-1,2-Dichloroethene
Concen: 17.039 ug/l
RT: 3.099 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Tgt Ion: 96 Resp: 17885
Ion Ratio Lower Upper
96 100
61 137.2 112.9 169.3
98 66.1 49.5 74.3



#22
Diisopropyl ether
Concen: 18.450 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion: 45 Resp: 66827
Ion Ratio Lower Upper
45 100
43 73.1 60.6 91.0
87 29.4 22.0 33.0
59 11.6 9.0 13.4









#23
 Vinyl Acetate
 Concen: 92.088 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: VX026735.D
 Acq: 02 Feb 2022 14:37




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0202WBSD03


Tgt Ion: 43 Resp: 274561
 Ion Ratio Lower Upper
 43 100
 86 10.9 8.1 12.1



#24
 1,1-Dichloroethane
 Concen: 17.561 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. -0.000 min
 Lab File: VX026735.D
 Acq: 02 Feb 2022 14:37

Tgt Ion: 63 Resp: 34351
 Ion Ratio Lower Upper
 63 100
 98 7.5 3.7 11.1
 100 4.7 2.4 7.2









#25
2-Butanone
Concen: 86.162 ug/l
RT: 4.562 min Scan# 51
Delta R.T. -0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Tgt Ion: 43 Resp: 74110
Ion Ratio Lower Upper
43 100
72 25.8 20.4 30.6



#26
2,2-Dichloropropane
Concen: 17.416 ug/l
RT: 4.477 min Scan# 556
Delta R.T. -0.006 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion: 77 Resp: 25253
Ion Ratio Lower Upper
77 100
97 22.8 11.0 33.0









#27
 cis-1,2-Dichloroethene
 Concen: 18.778 ug/l
 RT: 4.495 min Scan# 51
 Delta R.T. 0.000 min
 Lab File: VX026735.D
 Acq: 02 Feb 2022 14:37




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0202WBSD03


Tgt Ion: 96 Resp: 22496
 Ion Ratio Lower Upper
 96 100
 61 140.5 0.0 296.6
 98 64.1 0.0 128.8



#28
 Bromochloromethane
 Concen: 17.973 ug/l
 RT: 4.897 min Scan# 625
 Delta R.T. -0.006 min
 Lab File: VX026735.D
 Acq: 02 Feb 2022 14:37

Tgt Ion: 49 Resp: 14436
 Ion Ratio Lower Upper
 49 100
 129 1.6 0.0 4.2
 130 78.1 62.4 93.6





#29

Tetrahydrofuran

Concen: 95.865 ug/l

RT: 5.007 min Scan# 644

Delta R.T. -0.006 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37

Instrument :

MSVOA_X

ClientSampleId :

VX0202WBSD03





Tgt Ion: 42 Resp: 48972

Ion Ratio Lower Upper

42 100

72 43.1 34.6 52.0

71 39.8 31.8 47.8



#30

Chloroform

Concen: 18.211 ug/l

RT: 5.099 min Scan# 658

Delta R.T. 0.000 min

Lab File: VX026735.D




Acq: 02 Feb 2022 14:37


Tgt Ion: 83 Resp: 37014

Ion Ratio Lower Upper

83 100


85 64.8 51.6 77.4








#31
Cyclohexane
Concen: 15.718 ug/l
RT: 5.470 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03






Tgt Ion:	56	Resp:	30067
Ion	Ratio	Lower	Upper
56	100		
69	31.0	25.8	38.6
84	88.5	68.9	103.3



#32
1,1,1-Trichloroethane
Concen: 17.474 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion:	97	Resp:	31240
Ion	Ratio	Lower	Upper
97	100		
99	65.1	51.3	76.9
61	44.3	35.0	52.6





#33

1,2-Dichloroethane-d4

Concen: 43.747 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37

Instrument :

MSVOA_X

ClientSampleId :





VX0202WBSD03

Tgt Ion: 65 Resp: 56737

Ion Ratio Lower Upper

65 100

67 52.0 0.0 100.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. 0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37




Tgt Ion: 114 Resp: 149714


Ion Ratio Lower Upper

114 100

63 20.2 0.0 42.8

88 16.0 0.0 33.0









#35
Dibromofluoromethane
Concen: 47.076 ug/l
RT: 5.385 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Tgt Ion:113 Resp: 45513
Ion Ratio Lower Upper
113 100
111 101.8 81.2 121.8
192 19.8 15.0 22.4



#36
1,1-Dichloropropene
Concen: 16.883 ug/l
RT: 5.696 min Scan# 756
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion: 75 Resp: 25768
Ion Ratio Lower Upper
75 100
110 36.0 17.2 51.6
77 32.0 24.6 36.8





#37

Ethyl Acetate

Concen: 18.677 ug/l

RT: 4.715 min Scan# 596

Delta R.T. -0.006 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37





Instrument :

MSVOA_X

ClientSampleId :

VX0202WBSD03

Tgt Ion: 43	Resp: 28836
Ion Ratio	Lower Upper
43 100	
61 13.9	10.6 16.0
70 11.4	8.5 12.7



#38

Carbon Tetrachloride

Concen: 17.245 ug/l




RT: 5.678 min Scan# 753


Delta R.T. -0.006 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37

Tgt Ion: 117	Resp: 27073
Ion Ratio	Lower Upper
117 100	
119 93.5	74.7 112.1
121 29.8	25.4 38.0





#39

Methylcyclohexane

Concen: 17.139 ug/l

RT: 7.385 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37

Instrument :

MSVOA_X

ClientSampleId :

VX0202WBSD03





Tgt Ion: 83 Resp: 31553

Ion Ratio Lower Upper

83 100

55 79.5 63.6 95.4

98 50.1 38.4 57.6



#40

Benzene

Concen: 18.523 ug/l

RT: 6.044 min Scan# 813

Delta R.T. 0.000 min

Lab File: VX026735.D




Acq: 02 Feb 2022 14:37


Tgt Ion: 78 Resp: 78140

Ion Ratio Lower Upper

78 100

77 23.2 18.9 28.3









#41
Methacrylonitrile
Concen: 19.254 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Tgt Ion: 41	Resp: 16839
Ion Ratio	Lower Upper
41 100	
39 55.1	45.4 68.0
67 70.9	53.2 79.8
52 28.8	23.0 34.6



#42
1,2-Dichloroethane
Concen: 17.669 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion: 62	Resp: 29163
Ion Ratio	Lower Upper
62 100	
98 9.8	0.0 17.8





#43

Isopropyl Acetate

Concen: 18.850 ug/l

RT: 6.342 min Scan# 862

Delta R.T. 0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37

Instrument :

MSVOA_X

ClientSampleId :

VX0202WBSD03





Tgt Ion: 43 Resp: 47012

Ion Ratio Lower Upper

43 100

61 19.9 15.4 23.0

87 13.3 10.2 15.2



#44

Trichloroethene

Concen: 19.336 ug/l

RT: 7.129 min Scan# 991

Delta R.T. 0.000 min

Lab File: VX026735.D




Acq: 02 Feb 2022 14:37

Tgt Ion:130 Resp: 23146

Ion Ratio Lower Upper

130 100

95 100.5 0.0 202.6





#45

1,2-Dichloropropane

Concen: 19.357 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. 0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37

Instrument :

MSVOA_X

ClientSampleId :





VX0202WBSD03

Tgt Ion: 63 Resp: 20383

Ion Ratio Lower Upper

63 100

65 32.3 23.8 35.6



#46

Dibromomethane

Concen: 18.867 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37




Tgt Ion: 93 Resp: 14005


Ion Ratio Lower Upper

93 100

95 83.6 66.1 99.1

174 101.8 78.7 118.1









#47
Bromodichloromethane
Concen: 18.497 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Tgt Ion: 83	Resp: 27742
Ion Ratio	Lower Upper
83 100	
85 67.6	51.4 77.0
127 8.8	6.9 10.3



#48
Methyl methacrylate
Concen: 18.406 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion: 41	Resp: 22749
Ion Ratio	Lower Upper
41 100	
69 83.6	62.2 93.4
39 56.1	45.4 68.0





#49

1,4-Dioxane

Concen: 350.198 ug/l

RT: 7.683 min Scan# 1

Delta R.T. -0.012 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37

Instrument :

MSVOA_X

ClientSampleId :

VX0202WBSD03





Tgt Ion: 88 Resp: 7103

Ion Ratio Lower Upper

88 100

43 37.0 29.9 44.9

58 73.2 57.8 86.6



#50

Toluene-d8

Concen: 47.549 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.006 min

Lab File: VX026735.D




Acq: 02 Feb 2022 14:37


Tgt Ion: 98 Resp: 165657

Ion Ratio Lower Upper

98 100

100 65.7 53.2 79.8









#51
4-Methyl-2-Pentanone
Concen: 99.699 ug/l
RT: 8.574 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Tgt Ion: 43 Resp: 147479
Ion Ratio Lower Upper
43 100
58 39.1 30.5 45.7



#52
Toluene
Concen: 19.048 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion: 92 Resp: 49276
Ion Ratio Lower Upper
92 100
91 169.6 135.7 203.5





#53

t-1,3-Dichloropropene

Concen: 17.876 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.006 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37

Instrument :

MSVOA_X

ClientSampleId :





VX0202WBSD03

Tgt Ion: 75 Resp: 26649

Ion Ratio Lower Upper

75 100

77 31.5 25.7 38.5



#54

cis-1,3-Dichloropropene

Concen: 19.186 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37




Tgt Ion: 75 Resp: 30907


Ion Ratio Lower Upper

75 100

77 33.2 26.2 39.2

39 49.8 43.5 65.3





#55

1,1,2-Trichloroethane

Concen: 20.271 ug/l

RT: 9.153 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37

Instrument :

MSVOA_X

ClientSampleId :

VX0202WBSD03

Tgt Ion: 97 Resp: 20464





Ion Ratio Lower Upper

97 100

83 87.2 68.4 102.6

85 52.9 44.0 66.0

99 66.6 49.8 74.6



#56

Ethyl methacrylate

Concen: 20.367 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37




Tgt Ion: 69 Resp: 31510


Ion Ratio Lower Upper

69 100

41 66.9 57.2 85.8

39 34.7 30.9 46.3





#57

1,3-Dichloropropane

Concen: 19.791 ug/l

RT: 9.311 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37

Instrument :

MSVOA_X

ClientSampleId :





VX0202WBSD03

Tgt Ion: 76 Resp: 34638

Ion Ratio Lower Upper

76 100

78 32.0 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 90.614 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX026735.D




Acq: 02 Feb 2022 14:37


Tgt Ion: 63 Resp: 73510

Ion Ratio Lower Upper

63 100

106 29.8 23.0 34.6





#59

2-Hexanone

Concen: 94.130 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37

Instrument :

MSVOA_X

ClientSampleId :





VX0202WBSD03

Tgt Ion: 43 Resp: 106572

Ion Ratio Lower Upper

43 100

58 54.4 26.4 79.2



#60

Dibromochloromethane

Concen: 19.620 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. 0.000 min

Lab File: VX026735.D




Acq: 02 Feb 2022 14:37

Tgt Ion: 129 Resp: 20640

Ion Ratio Lower Upper

129 100

127 79.6 38.5 115.5









#61
1,2-Dibromoethane
Concen: 20.008 ug/l
RT: 9.610 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Tgt Ion:107 Resp: 21047
Ion Ratio Lower Upper
107 100
109 92.5 75.5 113.3



#62
4-Bromofluorobenzene
Concen: 48.031 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion: 95 Resp: 59445
Ion Ratio Lower Upper
95 100
174 75.3 0.0 148.6
176 72.7 0.0 141.4








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Tgt Ion:117 Resp: 128688
Ion Ratio Lower Upper
117 100
82 54.6 45.2 67.8
119 32.5 26.2 39.2



#64
Tetrachloroethene
Concen: 20.615 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion:164 Resp: 25061
Ion Ratio Lower Upper
164 100
166 124.4 101.8 152.6
129 95.8 83.3 124.9
131 90.0 78.0 117.0









#65
Chlorobenzene
Concen: 19.412 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Tgt Ion:112 Resp: 52607
Ion Ratio Lower Upper
112 100
114 31.7 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 19.538 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion:131 Resp: 19674
Ion Ratio Lower Upper
131 100
133 95.7 47.9 143.8
119 63.4 31.7 95.1





#67

Ethyl Benzene

Concen: 18.808 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. -0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37

Instrument :

MSVOA_X

ClientSampleId :





VX0202WBSD03

Tgt Ion: 91 Resp: 92623

Ion Ratio Lower Upper

91 100

106 32.3 24.7 37.1



#68

m/p-Xylenes

Concen: 38.462 ug/l

RT: 10.305 min Scan# 1512

Delta R.T. 0.000 min

Lab File: VX026735.D




Acq: 02 Feb 2022 14:37


Tgt Ion:106 Resp: 70884

Ion Ratio Lower Upper

106 100

91 202.1 165.3 247.9









#69
o-Xylene
Concen: 19.455 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Tgt Ion:106 Resp: 34597
Ion Ratio Lower Upper
106 100
91 214.6 109.7 329.0



#70
Styrene
Concen: 19.813 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion:104 Resp: 57492
Ion Ratio Lower Upper
104 100
78 50.7 42.5 63.7
103 54.2 43.3 64.9









#71
Bromoform
Concen: 18.685 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Tgt Ion:173 Resp: 13265
Ion Ratio Lower Upper
173 100
175 47.7 24.3 72.9
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion:152 Resp: 59140
Ion Ratio Lower Upper
152 100
115 70.1 44.2 132.6
150 165.1 0.0 348.2









#73
Isopropylbenzene
Concen: 18.902 ug/l
RT: 10.963 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Tgt Ion:105 Resp: 91629
Ion Ratio Lower Upper
105 100
120 26.4 13.1 39.1



#74
N-ethyl acetate
Concen: 19.058 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion: 43 Resp: 35666
Ion Ratio Lower Upper
43 100
70 42.0 32.9 49.3
55 24.4 18.3 27.5
61 23.8 18.9 28.3





#75

1,1,2,2-Tetrachloroethane

Concen: 19.337 ug/l

RT: 11.213 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37

Instrument :

MSVOA_X

ClientSampleId :

VX0202WBSD03





Tgt Ion: 83 Resp: 25160

Ion Ratio Lower Upper

83 100

131 10.4 5.3 15.9

85 63.8 32.0 96.2



#76

1,2,3-Trichloropropane

Concen: 22.982 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. -0.006 min

Lab File: VX026735.D




Acq: 02 Feb 2022 14:37


Tgt Ion: 75 Resp: 29925

Ion Ratio Lower Upper

75 100

77 32.7 18.3 54.8









#77
Bromobenzene
Concen: 19.389 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Tgt Ion:156 Resp: 21533
Ion Ratio Lower Upper
156 100
77 159.0 84.5 253.4
158 96.1 48.1 144.4



#78
n-propylbenzene
Concen: 18.720 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion: 91 Resp: 103487
Ion Ratio Lower Upper
91 100
120 23.0 11.6 34.7









#79
2-Chlorotoluene
Concen: 18.453 ug/l
RT: 11.366 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Tgt Ion: 91 Resp: 61859
Ion Ratio Lower Upper
91 100
126 33.5 16.4 49.2



#80
1,3,5-Trimethylbenzene
Concen: 18.973 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion: 105 Resp: 75826
Ion Ratio Lower Upper
105 100
120 48.5 24.0 72.0









#81
trans-1,4-Dichloro-2-butene
Concen: 19.841 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37




Instrument :
MSVOA_X
ClientSampleId :
VX0202WBSD03


Tgt Ion: 75 Resp: 6815
Ion Ratio Lower Upper
75 100
53 119.3 94.0 141.0
89 46.3 37.0 55.4



#82
4-Chlorotoluene
Concen: 18.487 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX026735.D
Acq: 02 Feb 2022 14:37

Tgt Ion: 91 Resp: 71223
Ion Ratio Lower Upper
91 100
126 29.1 14.1 42.4









#83
 tert-Butylbenzene
 Concen: 18.971 ug/l
 RT: 11.713 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX026735.D
 Acq: 02 Feb 2022 14:37




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0202WBSD03


Tgt Ion:119 Resp: 70778
 Ion Ratio Lower Upper
 119 100
 91 62.7 31.4 94.0
 134 23.8 11.8 35.3



#84
 1,2,4-Trimethylbenzene
 Concen: 18.831 ug/l
 RT: 11.750 min Scan# 1749
 Delta R.T. -0.000 min
 Lab File: VX026735.D
 Acq: 02 Feb 2022 14:37

Tgt Ion:105 Resp: 75273
 Ion Ratio Lower Upper
 105 100
 120 44.7 21.9 65.5









#85
 sec-Butylbenzene
 Concen: 19.242 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX026735.D
 Acq: 02 Feb 2022 14:37




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0202WBSD03


Tgt Ion:105 Resp: 92859
 Ion Ratio Lower Upper
 105 100
 134 20.5 10.1 30.1



#86
 p-Isopropyltoluene
 Concen: 19.217 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX026735.D
 Acq: 02 Feb 2022 14:37

Tgt Ion:119 Resp: 77804
 Ion Ratio Lower Upper
 119 100
 134 26.4 13.0 39.0
 91 24.1 12.6 37.6





#87

1,3-Dichlorobenzene

Concen: 18.996 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37

Instrument :

MSVOA_X

ClientSampleId :

VX0202WBSD03





Tgt Ion:146 Resp: 39543

Ion Ratio Lower Upper

146 100

111 41.4 21.3 63.7

148 63.4 31.6 95.0



#88

1,4-Dichlorobenzene

Concen: 18.679 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37




Tgt Ion:146 Resp: 39355


Ion Ratio Lower Upper

146 100

111 41.7 20.2 60.6

148 63.7 31.4 94.0





#89

n-Butylbenzene

Concen: 18.928 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37

Instrument :

MSVOA_X

ClientSampleId :

VX0202WBSD03





Tgt Ion: 91 Resp: 65764

Ion Ratio Lower Upper

91 100

92 54.2 27.2 81.6

134 27.1 13.1 39.1



#90

Hexachloroethane

Concen: 17.954 ug/l

RT: 12.542 min Scan# 1879

Delta R.T. 0.000 min

Lab File: VX026735.D




Acq: 02 Feb 2022 14:37


Tgt Ion: 117 Resp: 11145

Ion Ratio Lower Upper

117 100

201 76.1 35.6 106.8









#91
 1,2-Dichlorobenzene
 Concen: 19.728 ug/l
 RT: 12.335 min Scan# 1845
 Delta R.T. 0.000 min
 Lab File: VX026735.D
 Acq: 02 Feb 2022 14:37




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0202WBSD03


Tgt Ion:146 Resp: 38808
 Ion Ratio Lower Upper
 146 100
 111 42.5 21.8 65.3
 148 64.5 32.0 96.2



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 17.442 ug/l
 RT: 12.945 min Scan# 1945
 Delta R.T. 0.000 min
 Lab File: VX026735.D
 Acq: 02 Feb 2022 14:37

Tgt Ion: 75 Resp: 5485
 Ion Ratio Lower Upper
 75 100
 155 85.1 40.0 119.9
 157 109.9 51.6 154.8





#93

1,2,4-Trichlorobenzene

Concen: 19.443 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. -0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37

Instrument :

MSVOA_X

ClientSampleId :

VX0202WBSD03





Tgt Ion:180 Resp: 24522

Ion Ratio Lower Upper

180 100

182 94.3 46.7 140.0

145 31.9 16.3 48.8



#94

Hexachlorobutadiene

Concen: 20.047 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37




Tgt Ion:225 Resp: 10673


Ion Ratio Lower Upper

225 100

223 61.7 32.0 96.0

227 61.7 32.3 96.9





#95

Naphthalene

Concen: 19.834 ug/l

RT: 13.780 min Scan# 2081

Delta R.T. 0.006 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37

Instrument :

MSVOA_X

ClientSampleId :

VX0202WBSD03





Tgt Ion:128 Resp: 83436

Ion Ratio Lower Upper

128 100

127 12.8 10.3 15.5

129 10.7 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 20.091 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. -0.000 min

Lab File: VX026735.D

Acq: 02 Feb 2022 14:37

Tgt Ion:180 Resp: 24920

Ion Ratio Lower Upper

180 100

182 93.9 47.2 141.6

145 34.0 17.5 52.5

