

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020222\
 Data File : VX026732.D
 Acq On : 02 Feb 2022 13:10
 Operator : JC/MD
 Sample : N1301-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 55-TH-ST-34

Quant Time: Feb 03 03:03:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011022W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jan 10 12:50:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	16933	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	92537	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	81125	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	38058	27.951	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	93.167%
60) 4-Bromofluorobenzene	11.079	95	37358	27.593	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.967%
63) Toluene-d8	8.653	98	108358	29.075	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	96.933%
Target Compounds						
15) Acetone	2.386	58	708	4.238	ug/l	Qvalue 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020222\
Data File : VX026732.D
Acq On : 02 Feb 2022 13:10
Operator : JC/MD
Sample : N1301-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
55-TH-ST-34

Quant Time: Feb 03 03:03:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011022W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Jan 10 12:50:55 2022
Response via : Initial Calibration

