

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020222\
 Data File : VX026738.D
 Acq On : 02 Feb 2022 15:47
 Operator : JC/MD
 Sample : N1380-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MID0222

Quant Time: Feb 03 03:07:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	126126	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	208177	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186381	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80287	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	78409	44.503	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.000%
35) Dibromofluoromethane	5.391	113	64668	48.104	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.200%
50) Toluene-d8	8.653	98	243849	50.336	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.680%
62) 4-Bromofluorobenzene	11.085	95	88221	51.264	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.520%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.971	59	6160	25.133	ug/l	100
16) Acetone	2.386	43	1732	1.928	ug/l	97
52) Toluene	8.726	92	1047	0.291	ug/l	91
95) Naphthalene	13.780	128	2616	0.458	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020222\
Data File : VX026738.D
Acq On : 02 Feb 2022 15:47
Operator : JC/MD
Sample : N1380-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MID0222

Quant Time: Feb 03 03:07:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

