

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020222\
 Data File : VX026744.D
 Acq On : 02 Feb 2022 18:06
 Operator : JC/MD
 Sample : PB142421ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142421ZHE#03

Quant Time: Feb 03 03:08:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	121199	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	206330	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	178190	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78082	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	77907	46.016	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.040%
35) Dibromofluoromethane	5.391	113	64455	48.375	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.740%
50) Toluene-d8	8.653	98	237182	49.398	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.800%
62) 4-Bromofluorobenzene	11.079	95	83773	49.115	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.220%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	8615	9.978	ug/l	99
18) Methyl Acetate	2.703	43	2051	1.195	ug/l #	85
20) Methylene Chloride	2.794	84	3094	2.169	ug/l	90
43) Isopropyl Acetate	6.348	43	22902	6.663	ug/l	100
80) 1,3,5-Trimethylbenzene	11.457	105	1673	0.317	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020222\
Data File : VX026744.D
Acq On : 02 Feb 2022 18:06
Operator : JC/MD
Sample : PB142421ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142421ZHE#03

Quant Time: Feb 03 03:08:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

