

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020222\
 Data File : VX026746.D
 Acq On : 02 Feb 2022 18:53
 Operator : JC/MD
 Sample : PB142421ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142421ZHE#05

Quant Time: Feb 03 03:09:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	125986	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	213979	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186444	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83270	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79695	45.283	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.560%
35) Dibromofluoromethane	5.391	113	66231	47.931	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.860%
50) Toluene-d8	8.653	98	248289	49.863	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.720%
62) 4-Bromofluorobenzene	11.079	95	88814	50.209	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.420%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	8241	9.182	ug/l	97
18) Methyl Acetate	2.709	43	2392	1.341	ug/l	91
20) Methylene Chloride	2.794	84	3203	2.160	ug/l	89
43) Isopropyl Acetate	6.342	43	22246	6.241	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	1597	0.284	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020222\
Data File : VX026746.D
Acq On : 02 Feb 2022 18:53
Operator : JC/MD
Sample : PB142421ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142421ZHE#05

Quant Time: Feb 03 03:09:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

