

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020222\
 Data File : VX026747.D
 Acq On : 02 Feb 2022 19:16
 Operator : JC/MD
 Sample : PB142421ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142421ZHE#06

Quant Time: Feb 03 03:09:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	121316	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	205073	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182871	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81234	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77584	45.781	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	91.560%	
35) Dibromofluoromethane	5.391	113	63808	48.183	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	96.360%	
50) Toluene-d8	8.653	98	238619	50.002	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	100.000%	
62) 4-Bromofluorobenzene	11.079	95	87713	51.740	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	103.480%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	8947	10.353	ug/l #	89
18) Methyl Acetate	2.709	43	2383	1.387	ug/l	92
20) Methylene Chloride	2.794	84	2981	2.088	ug/l	96
25) 2-Butanone	4.574	43	1707	1.519	ug/l #	90
43) Isopropyl Acetate	6.349	43	22116	6.474	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	1678	0.306	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020222\
Data File : VX026747.D
Acq On : 02 Feb 2022 19:16
Operator : JC/MD
Sample : PB142421ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142421ZHE#06

Quant Time: Feb 03 03:09:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

