

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020222\
 Data File : VX026749.D
 Acq On : 02 Feb 2022 20:02
 Operator : JC/MD
 Sample : N1380-03DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 INF0222DL

Quant Time: Feb 03 03:09:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	121130	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	201715	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78194	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	74965	44.303	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.600%
35) Dibromofluoromethane	5.391	113	63038	48.393	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.780%
50) Toluene-d8	8.653	98	234796	50.020	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.040%
62) 4-Bromofluorobenzene	11.079	95	85431	51.233	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.460%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	911	1.056	ug/l	98
31) Cyclohexane	5.477	56	8050	3.225	ug/l	97
39) Methylcyclohexane	7.385	83	3873	1.561	ug/l	94
40) Benzene	6.044	78	132286	23.274	ug/l	97
52) Toluene	8.720	92	17461	5.010	ug/l	97
67) Ethyl Benzene	10.195	91	80452	11.533	ug/l	98
68) m/p-Xylenes	10.305	106	105078	40.253	ug/l	99
69) o-Xylene	10.647	106	14899	5.915	ug/l	96
73) Isopropylbenzene	10.964	105	8058	1.257	ug/l	97
78) n-propylbenzene	11.305	91	11971	1.638	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	17672	3.344	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	77304	14.627	ug/l	99
95) Naphthalene	13.774	128	32260	5.800	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020222\
Data File : VX026749.D
Acq On : 02 Feb 2022 20:02
Operator : JC/MD
Sample : N1380-03DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
INF0222DL

Quant Time: Feb 03 03:09:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

