

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020224\
 Data File : VX040086.D
 Acq On : 02 Feb 2024 14:53
 Operator : JC/MD
 Sample : P1343-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-GW-620-622

Quant Time: Feb 03 02:16:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	90771	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	163307	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	155422	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76897	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	66569	45.569	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.140%
35) Dibromofluoromethane	5.385	113	51304	46.150	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.300%
50) Toluene-d8	8.647	98	199479	47.982	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.960%
62) 4-Bromofluorobenzene	11.079	95	74148	46.319	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.640%
Target Compounds						
					Qvalue	
67) Ethyl Benzene	10.195	91	3530	0.612	ug/l	100
78) n-propylbenzene	11.305	91	3929	0.578	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	1092	0.230	ug/l	88
84) 1,2,4-Trimethylbenzene	11.750	105	6502	1.373	ug/l	95
95) Naphthalene	13.780	128	5335	1.014	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020224\
Data File : VX040086.D
Acq On : 02 Feb 2024 14:53
Operator : JC/MD
Sample : P1343-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-GW-620-622

Quant Time: Feb 03 02:16:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

