

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020224\
 Data File : VX040088.D
 Acq On : 02 Feb 2024 15:39
 Operator : JC/MD
 Sample : P1343-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-GW-680-682

Quant Time: Feb 03 02:17:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	82550	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	147798	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	123505	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	63791	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	62120	46.758	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.520%
35) Dibromofluoromethane	5.385	113	48485	48.190	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.380%
50) Toluene-d8	8.647	98	173190	46.030	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.060%
62) 4-Bromofluorobenzene	11.079	95	66666	46.015	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.020%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1638	2.631	ug/l	95
67) Ethyl Benzene	10.201	91	1365	0.298	ug/l #	55
78) n-propylbenzene	11.311	91	1694	0.301	ug/l	100
84) 1,2,4-Trimethylbenzene	11.756	105	2236	0.569	ug/l	98
95) Naphthalene	13.780	128	1957	0.448	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020224\
Data File : VX040088.D
Acq On : 02 Feb 2024 15:39
Operator : JC/MD
Sample : P1343-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-GW-680-682

Quant Time: Feb 03 02:17:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

