

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020224\
 Data File : VX040090.D
 Acq On : 02 Feb 2024 16:25
 Operator : JC/MD
 Sample : P1343-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-TB-20240129

Quant Time: Feb 03 02:18:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	80988	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	141233	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	137860	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57552	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61553	47.225	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.440%
35) Dibromofluoromethane	5.391	113	48415	50.358	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.720%
50) Toluene-d8	8.646	98	176682	49.140	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.280%
62) 4-Bromofluorobenzene	11.079	95	68137	49.216	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020224\
Data File : VX040090.D
Acq On : 02 Feb 2024 16:25
Operator : JC/MD
Sample : P1343-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-TB-20240129

Quant Time: Feb 03 02:18:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

