

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020224\
 Data File : VX040092.D
 Acq On : 02 Feb 2024 17:11
 Operator : JC/MD
 Sample : P1344-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-191-EB-20240131

Quant Time: Feb 03 02:19:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	75800	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	139422	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	138097	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	66811	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	59638	48.887	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.780%		
35) Dibromofluoromethane	5.385	113	45883	48.344	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.680%		
50) Toluene-d8	8.647	98	174920	49.282	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.560%		
62) 4-Bromofluorobenzene	11.079	95	67128	49.117	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	98.240%		
Target Compounds						
29) Tetrahydrofuran	5.013	42	6080	12.035	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020224\
Data File : VX040092.D
Acq On : 02 Feb 2024 17:11
Operator : JC/MD
Sample : P1344-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-191-EB-20240131

Quant Time: Feb 03 02:19:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

