

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020224\
 Data File : VX040096.D
 Acq On : 02 Feb 2024 18:43
 Operator : JC/MD
 Sample : P1343-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-GW-585-587

Quant Time: Feb 03 02:20:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	75734	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	137838	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124952	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64284	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	56394	46.268	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.540%
35) Dibromofluoromethane	5.385	113	44823	47.770	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.540%
50) Toluene-d8	8.647	98	160906	45.855	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.700%#
62) 4-Bromofluorobenzene	11.079	95	63180	46.760	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.520%
Target Compounds						
17) Carbon Disulfide	2.508	76	7143	2.868	ug/l	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020224\
Data File : VX040096.D
Acq On : 02 Feb 2024 18:43
Operator : JC/MD
Sample : P1343-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-GW-585-587

Quant Time: Feb 03 02:20:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

