

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020320\
 Data File : VX014787.D
 Acq On : 03 Feb 2020 10:35
 Operator : JC/SP
 Sample : VX0203WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0203WBL01

Quant Time: Feb 03 11:27:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 22 12:35:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	430323	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	685602	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	622861	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	277993	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	257247	51.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.38%	
35) Dibromofluoromethane	5.48	113	215977	49.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.84%	
50) Toluene-d8	8.71	98	809049	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
62) 4-Bromofluorobenzene	11.13	95	272545	46.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.00%	
Target Compounds						
16) Acetone	2.43	43	6666	2.569	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020320\
Data File : VX014787.D
Acq On : 03 Feb 2020 10:35
Operator : JC/SP
Sample : VX0203WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBL01

Quant Time: Feb 03 11:27:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 22 12:35:15 2020
Response via : Initial Calibration

