

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020320\
 Data File : VX014795.D
 Acq On : 03 Feb 2020 14:40
 Operator : JC/SP
 Sample : L1294-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M1-1-4-COMP

Quant Time: Feb 10 10:11:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\624X011520W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jan 15 12:41:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.01	128	68920	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.85	114	367649	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	327400	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.06	65	153731	31.73	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	105.77%
60) 4-Bromofluorobenzene	11.13	95	145905	27.85	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	92.83%
63) Toluene-d8	8.71	98	438499	29.65	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	98.83%

Target Compounds

						Qvalue
15) Acetone	2.43	58	5365	6.79	ug/l	78
18) Methylene Chloride	2.85	84	11485	2.57	ug/l	95
45) 1,4-Dioxane	7.73	88	15037	150.71	ug/l	99
64) Chlorobenzene	10.14	112	211131	19.47	ug/l	97
70) Isopropylbenzene	11.01	105	7492	0.41	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020320\
Data File : VX014795.D
Acq On : 03 Feb 2020 14:40
Operator : JC/SP
Sample : L1294-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M1-1-4-COMP

Quant Time: Feb 10 10:11:54 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\624X011520W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jan 15 12:41:12 2020
Response via : Initial Calibration

