

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX020320\
 Data File : VX014786.D
 Acq On : 03 Feb 2020 09:57
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 03 11:28:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 22 12:35:15 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.598	0.625	-4.5	90	0.00
3 P	Chloromethane	0.676	0.630	6.8	90	0.00
4 C	Vinyl Chloride	0.731	0.732	-0.1#	91	0.00
5 T	Bromomethane	0.497	0.451	9.3	91	0.00
6 T	Chloroethane	0.458	0.412	10.0	92	0.00
7 T	Trichlorofluoromethane	0.852	0.827	2.9	97	0.00
8 T	Diethyl Ether	0.372	0.356	4.3	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.500	0.503	-0.6	100	0.00
10 T	Methyl Iodide	0.656	0.667	-1.7	92	0.00
11 T	Tert butyl alcohol	0.131	0.105	19.8	91	0.00
12 CM	1,1-Dichloroethene	0.497	0.476	4.2#	95	0.00
13 T	Acrolein	0.103	0.101	1.9	100	0.00
14 T	Allyl chloride	0.893	0.885	0.9	95	0.00
15 T	Acrylonitrile	0.303	0.289	4.6	96	0.00
16 T	Acetone	0.348	0.372	-6.9	116	0.00
17 T	Carbon Disulfide	1.358	1.253	7.7	90	0.00
18 T	Methyl Acetate	0.913	0.883	3.3	96	0.00
19 T	Methyl tert-butyl Ether	1.628	1.655	-1.7	99	0.00
20 T	Methylene Chloride	0.592	0.557	5.9	97	0.00
21 T	trans-1,2-Dichloroethene	0.547	0.515	5.9	95	0.00
22 T	Diisopropyl ether	1.777	1.836	-3.3	98	0.00
23 T	Vinyl Acetate	1.355	1.414	-4.4	102	0.00
24 P	1,1-Dichloroethane	0.965	0.964	0.1	97	0.00
25 T	2-Butanone	0.458	0.453	1.1	99	0.00
26 T	2,2-Dichloropropane	0.773	0.799	-3.4	101	0.00
27 T	cis-1,2-Dichloroethene	0.613	0.598	2.4	97	0.00
28 T	Bromochloromethane	0.424	0.387	8.7	97	0.00
29 T	Tetrahydrofuran	0.272	0.250	8.1	91	0.00
30 C	Chloroform	0.956	0.964	-0.8#	100	-0.01
31 T	Cyclohexane	0.908	0.865	4.7	92	0.00
32 T	1,1,1-Trichloroethane	0.808	0.813	-0.6	100	0.00
33 S	1,2-Dichloroethane-d4	0.584	0.577	1.2	100	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.315	0.316	-0.3	99	0.00
36 T	1,1-Dichloropropene	0.465	0.483	-3.9	97	0.00
37 T	Ethyl Acetate	0.532	0.519	2.4	94	0.00
38 T	Carbon Tetrachloride	0.444	0.474	-6.8	101	0.00
39 T	Methylcyclohexane	0.582	0.594	-2.1	94	0.00
40 TM	Benzene	1.436	1.453	-1.2	96	0.00
41 T	Methacrylonitrile	0.319	0.295	7.5	95	0.00
42 TM	1,2-Dichloroethane	0.496	0.522	-5.2	100	0.00
43 T	Isopropyl Acetate	0.871	0.863	0.9	95	0.00
44 TM	Trichloroethene	0.405	0.403	0.5	96	0.00
45 C	1,2-Dichloropropane	0.370	0.388	-4.9#	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX020320\
 Data File : VX014786.D
 Acq On : 03 Feb 2020 09:57
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 03 11:28:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 22 12:35:15 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.240	0.253	-5.4	100	0.00
47 T	Bromodichloromethane	0.470	0.511	-8.7	101	0.00
48 T	Methyl methacrylate	0.416	0.432	-3.8	95	0.00
49 T	1,4-Dioxane	0.009	0.007	22.2#	88	0.00
50 S	Toluene-d8	1.213	1.199	1.2	97	0.00
51 T	4-Methyl-2-Pentanone	0.550	0.540	1.8	95	0.00
52 CM	Toluene	0.882	0.921	-4.4#	97	0.00
53 T	t-1,3-Dichloropropene	0.506	0.557	-10.1	100	0.00
54 T	cis-1,3-Dichloropropene	0.556	0.618	-11.2	100	0.00
55 T	1,1,2-Trichloroethane	0.361	0.379	-5.0	98	0.00
56 T	Ethyl methacrylate	0.544	0.581	-6.8	96	0.00
57 T	1,3-Dichloropropane	0.618	0.640	-3.6	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.182	0.199	-9.3	97	0.00
59 T	2-Hexanone	0.427	0.427	0.0	95	0.00
60 T	Dibromochloromethane	0.377	0.412	-9.3	102	0.00
61 T	1,2-Dibromoethane	0.382	0.397	-3.9	99	0.00
62 S	4-Bromofluorobenzene	0.432	0.426	1.4	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.495	0.470	5.1	88	0.00
65 PM	Chlorobenzene	1.069	1.098	-2.7	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.417	-5.8	102	0.00
67 C	Ethyl Benzene	1.826	1.950	-6.8#	99	0.00
68 T	m/p-Xylenes	0.692	0.747	-7.9	99	0.00
69 T	o-Xylene	0.668	0.711	-6.4	98	0.00
70 T	Styrene	1.120	1.247	-11.3	99	0.00
71 P	Bromoform	0.314	0.349	-11.1	104	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.516	3.738	-6.3	100	0.00
74 T	N-amyl acetate	1.541	1.585	-2.9	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.255	1.204	4.1	100	0.00
76 T	1,2,3-Trichloropropane	1.163	1.027	11.7	86	0.00
77 T	Bromobenzene	0.944	0.967	-2.4	100	0.00
78 T	n-propylbenzene	3.991	4.332	-8.5	101	0.00
79 T	2-Chlorotoluene	2.415	2.525	-4.6	100	0.00
80 T	1,3,5-Trimethylbenzene	2.918	3.169	-8.6	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.409	0.423	-3.4	99	0.00
82 T	4-Chlorotoluene	2.749	2.919	-6.2	100	0.00
83 T	tert-Butylbenzene	2.742	2.993	-9.2	106	0.00
84 T	1,2,4-Trimethylbenzene	2.927	3.168	-8.2	101	0.00
85 T	sec-Butylbenzene	3.359	3.681	-9.6	101	0.00
86 T	p-Isopropyltoluene	3.104	3.418	-10.1	102	0.00
87 T	1,3-Dichlorobenzene	1.697	1.736	-2.3	101	0.00
88 T	1,4-Dichlorobenzene	1.700	1.745	-2.6	102	0.00
89 T	n-Butylbenzene	2.704	3.008	-11.2	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX020320\
Data File : VX014786.D
Acq On : 03 Feb 2020 09:57
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 03 11:28:36 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 22 12:35:15 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.577	0.620	-7.5	106	0.00
91 T	1,2-Dichlorobenzene	1.686	1.728	-2.5	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.270	0.257	4.8	97	0.00
93 T	1,2,4-Trichlorobenzene	1.087	1.183	-8.8	102	0.00
94 T	Hexachlorobutadiene	0.559	0.598	-7.0	103	0.00
95 T	Naphthalene	3.127	3.418	-9.3	97	0.00
96 T	1,2,3-Trichlorobenzene	1.099	1.173	-6.7	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6