

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020321\
 Data File : VX020835.D
 Acq On : 03 Feb 2021 13:09
 Operator : JC/SP
 Sample : M1289-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-MW-2S(OFFSITE)

Quant Time: Feb 04 02:21:22 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	285199	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	488264	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	476092	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	203782	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	195682	50.30	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.60%	
35) Dibromofluoromethane	5.464	113	169192	50.59	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.18%	
50) Toluene-d8	8.695	98	600128	48.95	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 97.90%	
62) 4-Bromofluorobenzene	11.122	95	211858	44.49	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 88.98%	
Target Compounds						
						Qvalue
27) cis-1,2-Dichloroethene	4.562	96	8034	2.14	ug/l	87
30) Chloroform	5.160	83	1404	0.23	ug/l	87
44) Trichloroethene	7.177	130	1099	0.29	ug/l	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020321\
Data File : VX020835.D
Acq On : 03 Feb 2021 13:09
Operator : JC/SP
Sample : M1289-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-MW-2S(OFFSITE)

Quant Time: Feb 04 02:21:22 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

