

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020321\
 Data File : VX020839.D
 Acq On : 03 Feb 2021 14:43
 Operator : JC/SP
 Sample : M1289-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-DDC-5-PS

Quant Time: Feb 04 02:22:43 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	280002	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	473289	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	452040	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	180693	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	190811	49.96	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.92%
35) Dibromofluoromethane	5.464	113	165043	50.91	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.82%
50) Toluene-d8	8.696	98	584253	49.16	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.32%
62) 4-Bromofluorobenzene	11.122	95	196298	42.53	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	85.06%
Target Compounds						
16) Acetone	2.422	43	2286	1.58	ug/l	Qvalue 98
27) cis-1,2-Dichloroethene	4.562	96	23528	6.39	ug/l	89
30) Chloroform	5.178	83	2133	0.35	ug/l #	74
64) Tetrachloroethene	9.317	164	4506	1.30	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020321\
Data File : VX020839.D
Acq On : 03 Feb 2021 14:43
Operator : JC/SP
Sample : M1289-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-DDC-5-PS

Quant Time: Feb 04 02:22:43 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

